

AIMS

Asbestos in Materials Scheme

Dr S Pierdzig
CRB Analyse Service GmbH
Bahnhofstrasse 14
D-37181 Hardegsen

37181

GERMANY



Round 43 Report

Please find enclosed your results for round 43.

Regards

Delia Lomas-Fletcher
Senior Scientist
AIMS PT Manager

Tel: 01298 218553

Fax: 01298 218172

Email: proficiency.testing@hsl.gov.uk

Asbestos in Materials Individual Results Round 43

for Laboratory Number 1640

Printed on 29 Mar 2011


Your Performance Score:	9999	Your Performance Classification:	Unclassified
		Your Score for this round:	0

Sample 1	Your Results	Assigned Results	Median	Your Score
	Other No Asbestos	No Asbestos		0

Sample 2	Your Results	Assigned Results	Median	Your Score
	No Asbestos	No Asbestos		0

Sample 3	Your Results	Assigned Results	Median	Your Score
	Chrysotile	Chrysotile		0



Sample 4	Your Results	Assigned Results	Median	Your Score
	Amosite MMMF	Amosite		0



Asbestos in Materials Results Round 43

for Laboratory Number 1640

Printed on 29 Mar 2011



Quantitative Results for Sample 1

