

# PT REPORT

Number: RIAB88.48 - Final

Page 1 of 2

Description	RIAB - Ring Investigation Identification Asbestos in Bulk Materials.
Participant	CRB Analyse Service GmbH Bahnhofstrasse 14 37181 Hardegsen Germany
Method and evaluation	The PT was organised in accordance with document: Protocol RIAB. After obtaining the reference value, the samples have been shipped to the participants laboratory. The reported results of the participant are compared with the reference values. The results of the comparison are presented on page 2.
Policy statement:	The evaluation reports of VSL's proficiency testing schemes are provided for the purpose of communicating the proficiency demonstrated by participants on specific measurements and/or calibrations. The reports are intended to be used in support of demonstrating competence in measurement or calibration, fulfilling quality control requirements as stipulated in written standards on showing such competence, and claims of calibration and measurement capabilities.

Delft, 7 January 2022  
VSL B.V.

R.P. van Otterloo



# PT REPORT

Number: RIAB88.48 - Final  
Page 2 of 2

Period of ILC November 2021 – December 2021

Results &  
Evaluation

Fibre type	8801		8802	
	Results	Reference	Results	Reference
Chrysotile	< 0.1%	< 0.1%	10-15%	10-15%
Amosite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Crocidolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Tremolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Actinolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Anthophyllite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
	Points: 10		Points: 10	

Fibre type	8803		8804	
	Results	Reference	Results	Reference
Chrysotile	2-5%	5-10%	30-60%	> 60%
Amosite	15-30%	15-30%	< 0.1%	< 0.1%
Crocidolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Tremolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Actinolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Anthophyllite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
	Points: 10		Points: 10	

Results The result with a total score of 100 is satisfactory

# RIAB 88 results

