

Fluxana, Deutschland - Tonerde FLX-CRM 112

Veranstalter: FluXana GmbH & Co. KG, Borschelstr. 3, 47551 Bedburg-Hau

Ringversuchsmaterial: FLX-CRM 111 - Chrommagnetit, Germany

RV geschlossen: 2012 – 9

Literatur: Proficiency Test Report Fluxana FLX-CRM 111, 112 (CRB Laborcode = 23)

Hauptelemente [MA%]

	CRB	RV	1sRV	Z-Score
Al ₂ O ₃	80,04	79,813	0,536	0,41
CaO	0,16	0,147	0,027	0,49
Fe ₂ O ₃	0,39	0,326	0,074	0,87
K ₂ O	0,084	0,090	0,005	1,16
MgO	0,72	0,755	0,080	0,44
Na ₂ O	0,33	0,267	0,066	0,96
SiO ₂	11,90	12,164	0,254	1,04
P ₂ O ₅	0,082	0,074	0,012	0,71
TiO ₂	0,20	0,273	0,025	2,94
LOI	5,01	5,424	0,126	3,32

Spurenelemente [µg/g]

	CRB	RV	1sRV	Z-Score
Cr ₂ O ₃	0,021	0,017	0,005	0,64
Mn ₃ O ₄	0,027	0,024	0,010	0,23
WO ₃	0,013	0,037	0,024	1,02
Co ₃ O ₄	0,007	0,010	0,007	0,49
ZrO ₂	5,86	5,949	0,333	0,28
NiO	0,010	0,009	0,004	1,10
HfO ₂	0,122	0,097	0,029	0,84

Legende

CRB: Ergebnisse CRB – **RV:** Ergebnisse Ringversuch -- **1s-RV:** Standardabweichung Ringversuch

Z-Score: Differenz des Messwertes vom Mittelwert des Ringversuchs -- * Wert nicht zertifiziert

ZERTIFIKAT

FLX CRM-111-112 Ringversuch 2012

CRB Analyse Service GmbH
Bahnhofstraße 14
37181 Hardegsen

Das Testlabor nahm mit der Labornummer 23 an dem FLUXANA Ringversuch FLX CRM-111-112 mit Erfolg teil.


Probe: FLX CRM-111-112

Von 34 Parametern wurden 91% erfolgreich (d.h. nicht als Ausreißer detektiert) bestimmt:
Haupt- und Spurenelemente: Al₂O₃, CaO, Fe₂O₃, HfO₂, K₂O, MgO, NiO, SiO₂, SO₃,
TiO₂, ZrO₂, Co₃O₄, WO₃, Cr₂O₃, Mn₃O₄, P₂O₅, LOI

Die Durchführung dieses Ringversuchs und damit die Produktion dieses Referenzmaterials wurde in Übereinstimmung mit ISO Guide 34-2009, ISO Guide 31-2000 und ISO Guide 35-2006 ausgeführt.

Die Ergebnisse wurden in einem Bericht und Zertifikat zusammengefasst.

Bedburg-Hau, den 26.10.2012



Dr. Rainer Schramm

FLUXANA GmbH & Co. KG
Borschelstr. 3
47551 Bedburg-Hau

FLUXANA

New Certified Reference Materials

FLX-CRM 111, FLX-CRM 112



Proficiency Test Report

FLX-CRM 111, FLX-CRM 112

Introduction

X-ray fluorescence analysis is a widely used technique for the analysis of oxidic materials. Different ISO methods like e.g. 12677:2011 or 29581-2:2010 describe the use in detail.

However for the calibration of xrf instruments dedicated standard material is needed. As a world wide supplier for xrf laboratories FLUXANA has developed a number of services to support xrf users. One of these services is the production of new reference materials in combination with a proficiency test.

From 2011 FLUXANA has introduced its own quality management in agreement with ISO 17025.

The production of reference materials and the corresponding proficiency tests including all evaluations are performed in agreement with ISO 17043, ISO Guide 34-2009, ISO Guide 31-2000 and ISO Guide 35-2006.

Proficiency test

All laboratories which applied until 29.02.2012 for the participation of the proficiency test got their samples starting week 10 and sent in their results until 31.05.2012.

Further information

In the following evaluation report all laboratory data are listed. Also the used methods like XRF according ISO 12677, XRF preparation as fused bead, XRF preparation as pressed pellet, XRF as reconstitution method, ICP-OES, combustion with HF-IR (high frequency infrared) or others are specified. Laboratories which are working under ISO 17025 accreditation are highlighted. Under Remark additional information is given.

Certificate of Analysis

Based on this report a certificate of analysis is issued separately.

Outlier evaluation

There will be two outlier tests based on **Grubbs** and **z-score**.

However every outlier detected by the test was verified individually. Sometimes a value detected as outlier is included to guarantee a balance between different analytical methods. These values are marked as **'included'**. Real outliers which were excluded from the calculation of mean, standard deviation and uncertainty are marked as **'confirmed outliers'**.

Statistical Evaluation for a new RM (reference material)

All mentioned calculations are based on:

Reference materials – General and statistical principles for certification ISO Guide 35:2006.

Conformity assessment - General requirements for proficiency testing ISO 17043:2010.

Calculation of laboratory average

Each participant of the proficiency test must perform a number of single measurements and report with significant digits.

For each participant a laboratory average \bar{x} is calculated:

$$(1) \quad \bar{x} = \sum_{k=1}^p \frac{x}{p}$$

1 x Single measurement

p Number of single measurements

Calculation of total average

From all laboratory averages a total average $\bar{\bar{x}}$ is calculated:

$$(2) \quad \bar{\bar{x}} = \sum_{k=1}^n \frac{\bar{x}}{n}$$

n Number of participants

Calculation of standard deviation

From all laboratory averages the standard deviation s is calculated:

$$(3) \quad s = \sqrt{\sum_{k=1}^n (\bar{x} - \bar{\bar{x}})^2 / (n - 1)}$$

Test for outliers

From all laboratory averages the **z-score** z is calculated:

$$(4) \quad z = \left| \frac{(\bar{x} - \bar{x})}{s} \right|$$

An outlier test based on z-score is performed:

$z \leq 2,0$	indicates ‚satisfactory‘ performance = generates no signal
$2,0 < z < 3,0$	indicates ‚questionable‘ performance = generates a warning signal
$z \geq 3,0$	indicates ‚unsatisfactory‘ performance = generates an action signal

Parallel an outlier test based on Grubbs is performed:

$$(5) \quad PG = \left| \frac{(\bar{x} - \bar{x})}{s} \right|$$

PG test value

Based on table 1 a comparison value for the half width confidence interval is calculated for n:

$PG \leq VG$	indicates ‚satisfactory‘ performance = generates no signal
$PG > VG$	indicates ‚unsatisfactory‘ performance = generates an action signal

In case an outlier is detected the data will be taken out and all calculations according formulars 2,3,4,5 have to be repeated. A new test for outliers must be performed.

Calculation of the uncertainty

The uncertainty values are coming from the half width confidence interval C(95%). It is equal to:

$$(6) \quad C(95\%) = t * s / \sqrt{n}$$


t Student's value

where t is the appropriate Student's value, n the number of acceptable mean values and s the standard deviation.


Participating Laboratories

Bachema AG	Schlieren	Switzerland
BASF Coatings GmbH	Münster	Germany
CRB Analyse Service GmbH	Hardegsen	Germany
Deutsches Institut für Feuerfest und Keramik GmbH	Bonn	Germany
Dorfner Anzaplan	Hirschau	Germany
FLUXANA GmbH & Co.KG	Bedburg-Hau	Germany
Fundacion ITMA	Llanera-Asturias	Spain
Grothe Rohstoffe GmbH & Co. KG	Bückeberg	Germany
Holcim (Deutschland) AG	Sehnde	Germany
Hoppecke Batterien GmbH & Co. KG	Brilon-Hoppecke	Germany
HuK Umweltlabor GmbH	Wenden-Hünsborn	Germany
Imerys Minerals Ltd.	Cornwall	England
Instituto Nacional del Carbón (INCAR-CSIC)	Oviedo (Asturias)	Spain
Rockwool BV	JG Roermond	Netherlands
ThyssenKrupp Steel Europe AG	Duisburg	Germany
Vargön Alloys AB	Vargön	Sweden

Al2O3		FLX-CRM 111			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:28:20	z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean	z>3		n=13 VG=2,331	confirmed	
21	ISO 12677			3,942	4,232	x4,087		3,68	Outlier	x	
22	ISO 12677		reconstitution	4,726	4,443	4,585		0,50			
23	ISO 12677	YES		4,770	4,630	4,700		0,23			
24	ISO 12677			4,640	4,668	4,654		0,06			
25	XRF bead			4,700	4,440	4,570		0,60			
26	XRF pellet			3,832	3,706	x3,769		5,71	Outlier	x	
27	ISO 12677	YES		4,680	4,520	4,600		0,41			
28	ISO 12677	YES		4,670	4,730	4,700		0,23			
41	ICP-OES	YES	ISO 11885-E22	4,500		4,500		1,04			
29	ISO 12677	YES		4,650	4,722	4,686		0,14			
30	XRF bead			4,130	4,180	x4,155		3,24	Outlier	x	
31	ISO 12677	YES		0,837	1,061	x0,949		23,70	Outlier	x	
35	ISO 12677			4,784	4,714	4,749		0,55			
37	ICP-OES		Borax fusion	4,310		4,310		2,26	Included		
38	ISO 12677	YES		4,814	4,870	4,842		1,14			
39	XRF bead			4,780	4,930	4,855		1,22			
40	ISO 12677			4,936	4,815	4,876		1,35			
				n	13						
				Mean	4,664						
				Max	4,876						
				Min	4,310						
				Stdev s	0,157						
				C(95%)	0,095		C(95%)=t*s/SQR(n) t(13)=2,179				



CaO		FLX-CRM 111			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:27:51	z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean	z>3		n=15 VG=2,409	confirmed	
21	ISO 12677			1,975	1,888	1,932		2,12			
22	ISO 12677		reconstitution	1,871	1,830	x1,851		3,41	Outlier	x	
23	ISO 12677	YES		2,040	2,030	2,035		0,48			
24	ISO 12677			2,089	2,108	2,099		0,53			
25	XRF bead			2,050	2,060	2,055		0,16			
26	XRF pellet			2,851	2,828	x2,840		12,32	Outlier	x	
27	ISO 12677	YES		2,070	2,090	2,080		0,23			
28	ISO 12677	YES		2,120	2,140	2,130		1,03			
41	ICP-OES	YES	ISO 11885-E22	2,000		2,000		1,04			
29	ISO 12677	YES		2,015	2,050	2,033		0,52			
30	XRF bead			2,140	2,110	2,125		0,95			
31	ISO 12677	YES		2,141	2,122	2,132		1,05			
35	ISO 12677			2,138	2,123	2,131		1,04			
37	ICP-OES		aqua regia	1,960		1,960		1,68			
38	ISO 12677	YES		2,094	2,071	2,083		0,27			
39	XRF bead			2,080	2,140	2,110		0,71			
40	ISO 12677			2,069	2,087	2,078		0,20			
				n	15						
				Mean	2,065						
				Max	2,132						
				Min	1,932						
				Stdev s	0,063						
				C(95%)	0,035		C(95%)=t*s/SQR(n) t(15)=2,145				



Fe2O3		FLX-CRM 111			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:28:29		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean				z>3	n=14 VG=2,371	confirmed
21	ISO 12677			8,367	8,323	x8,345				7,58	Outlier	x
22	ISO 12677		reconstitution	9,482	9,457	9,470				0,44		
23	ISO 12677	YES		9,310	9,400	9,355				1,17		
24	ISO 12677			9,407	9,457	9,432				0,68		
25	XRF bead			9,620	9,690	9,655				0,73		
26	XRF pellet			9,483	9,818	9,651				0,71		
27	ISO 12677	YES		9,500	9,610	9,555				0,10		
28	ISO 12677	YES		8,770	8,800	x8,785				4,79	Outlier	x
41	ICP-OES	YES	ISO 11885-E22	8,400		x8,400				7,23	Outlier	x
29	ISO 12677	YES		9,768	9,656	9,712				1,10		
30	XRF bead			9,280	9,300	9,290				1,58		
31	ISO 12677	YES		9,555	9,522	9,539				0,01		
35	ISO 12677			9,715	9,659	9,687				0,94		
37	ICP-OES		Borax fusion	9,250		9,250				1,84		
38	ISO 12677	YES		9,699	9,607	9,653				0,72		
39	XRF bead			9,810	9,640	9,725				1,18		
40	ISO 12677			9,522	9,634	9,578				0,25		
						n	14					
						Mean	9,539					
						Max	9,725					
						Min	9,250					
						Stdev s	0,158					
						C(95%)	0,091					

C(95%)=t*s/SQR(n) t(14)=2,160

HfO2		FLX-CRM 111			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:10:27		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean				z>3	n=5 VG=1,672	confirmed
21	XRF pellet			0,001	0,001	0,001				0,96		
22												
23	ISO 12677	YES		0,001	0,001	0,001				0,15		
24	ISO 12677			<0,02	<0,02							
25												
26	XRF pellet			0,001	0,001	0,001				0,29		
27	ICP-OES	YES		0,001	0,002	0,001				0,54		
28	ISO 12677	YES		<0,01	<0,01							
41	ICP-OES	YES	ISO 11885-E22	<0,005								
29												
30												
31	ISO 12677	YES		0,007	0,007	x0,007				22,10	Outlier	x
35	ISO 12677			0,001	0,001	0,001				1,64		
37												
38	ISO 12677	YES		0,154	0,154	x0,154				567,29	Outlier	x
39												
40	ISO 12677											
						n	5					
						Mean	0,001					
						Max	0,001					
						Min	0,001					
						Stdev s	0,000					
						C(95%)	0,000					

C(95%)=t*s/SQR(n) t(5)=2,776

K2O		FLX-CRM 111			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:28:38		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			z>3	n=9 VG=2,110	confirmed	
21	ISO 12677			0,026	0,044	x0,035			4,53	Outlier	x	
22												
23	ISO 12677	YES		0,005	0,007	0,006						
24	ISO 12677			<0,03	<0,03							
25	XRF bead			<0,1	<0,1							
26	XRF pellet			0,013	0,007	0,010						
27	ICP-OES	YES		0,020	0,010	0,015						
28	ISO 12677	YES		0,020	0,020	0,020						
41	ICP-OES	YES	ISO 11885-E22	0,006		0,006						
29	ISO 12677	YES		0,001	0,001	0,001						
30	XRF bead			0,066	0,071	x0,069						
31	ISO 12677	YES		<0,0012	<0,0012							
35	ISO 12677			<0,0100	<0,0100							
37	ICP-OES		aqua regia	0,012		0,012						
38	ISO 12677	YES		<0,01	<0,01							
39	XRF bead			0,010	0,010	0,010						
40	ISO 12677			0,011	0,010	0,011						
						n	9					
						Mean	0,010					
						Max	0,020					
						Min	0,001					
						Stdev s	0,006					
						C(95%)	0,004					

C(95%)=t*s/SQR(n) t(9)=2,306

MgO		FLX-CRM 111			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:28:49		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			z>3	n=16 VG=2,443	confirmed	
21	ISO 12677			70,605	69,964	70,285						
22	ISO 12677		reconstitution	69,913	69,945	69,929						
23	ISO 12677	YES		70,080	70,070	70,075						
24	ISO 12677			70,970	71,284	71,127						
25	XRF bead			70,590	69,930	70,260						
26	XRF pellet			66,790	66,370	x66,580						
27	ISO 12677	YES		69,900	69,800	69,850						
28	ISO 12677	YES		71,300	70,900	71,100						
41	ICP-OES	YES	ISO 11885-E22	72,900		72,900						
29	ISO 12677	YES		69,669	70,284	69,977						
30	XRF bead			70,940	70,426	70,683						
31	ISO 12677	YES		72,780	72,650	72,715						
35	ISO 12677			68,605	68,200	68,403						
37	ICP-OES		aqua regia	67,840		67,840						
38	ISO 12677	YES		69,455	69,333	69,394						
39	XRF bead			69,010	68,650	68,830						
40	ISO 12677			69,883	69,917	69,900						
						n	16					
						Mean	70,204					
						Max	72,900					
						Min	67,840					
						Stdev s	1,347					
						C(95%)	0,717					

C(95%)=t*s/SQR(n) t(16)=2,131


Na2O		FLX-CRM 111			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:28:58		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			z>3	n=10 VG=2,176	confirmed	
21	ISO 12677			0,500	0,483	x0,492			22,51	Outlier	x	
22	ISO 12677		reconstitution	0,049	0,075	0,062			0,39			
23	ISO 12677	YES		<0,03	<0,03							
24												
25	ICP-OES			0,060	0,060	0,060			0,50			
26	XRF pellet			0,351	0,368	x0,360			15,48	Outlier	x	
27	ICP-OES	YES		0,070	0,080	0,075			0,30			
28	ISO 12677	YES		0,080	0,070	0,075			0,30			
41	ICP-OES	YES	ISO 11885-E22	0,050		0,050			1,03			
29	ISO 12677	YES		0,104	0,109	0,107			1,98			
30	XRF bead			0,277	0,289	x0,283			11,38	Outlier	x	
31	ISO 12677	YES		<0,014	<0,014							
35	ISO 12677			0,038	0,045	0,042			1,49			
37	ICP-OES		aqua regia	0,067		0,067			0,13			
38	ISO 12677	YES		0,072	0,062	0,067			0,13			
39	XRF bead			0,080	0,100	0,090			1,10			
40	ISO 12677			<0,1	<0,1							
				n	10							
				Mean	0,069							
				Max	0,107							
				Min	0,042							
				Stdev s	0,019							
				C(95%)	0,013							

C(95%)=t*s/SQR(n) t(10)=2,262


NiO		FLX-CRM 111			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:11:18		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			z>3	n=14 VG=2,371	confirmed	
21	XRF pellet			0,031	0,031	0,031			0,02			
22	ISO 12677		reconstitution	0,018	0,018	0,018			1,87			
23	ISO 12677	YES		0,030	0,031	0,031			0,05			
24	ISO 12677			0,033	0,033	0,033			0,30			
25												
26	XRF pellet			0,033	0,031	0,032			0,20			
27	ICP-OES	YES		0,017	0,017	0,017			1,97			
28	ISO 12677	YES		0,040	0,040	0,040			1,33			
41	ICP-OES	YES	ISO 11885-E22	0,040		0,040			1,33			
29	ISO 12677	YES		0,029	0,029	0,029			0,27			
30												
31	ISO 12677	YES		0,035	0,034	0,035			0,54			
35	ISO 12677			0,031	0,032	0,031			0,07			
37	ICP-OES		aqua regia	0,026		0,026			0,70			
38	ISO 12677	YES		0,031	0,031	0,031			0,02			
39												
40	ISO 12677			0,038	0,038	0,038			1,04			
				n	14							
				Mean	0,031							
				Max	0,040							
				Min	0,017							
				Stdev s	0,007							
				C(95%)	0,004							

C(95%)=t*s/SQR(n) t(14)=2,160

SiO2		FLX-CRM 111			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:29:06		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			z>3	n=14 VG=2,371	confirmed	
21	ISO 12677			1,694	2,177	x1,936			4,78	Outlier	x	
22	ISO 12677		reconstitution	1,311	1,299	1,305			0,84			
23	ISO 12677	YES		1,290	1,260	1,275			1,11			
24	ISO 12677			1,302	1,301	1,302			0,87			
25	XRF bead			1,310	1,390	1,350			0,44			
26	XRF pellet			2,654	2,552	x2,603			10,72	Outlier	x	
27	ISO 12677	YES		1,350	1,400	1,375			0,22			
28	ISO 12677	YES		1,610	1,630	1,620			1,97			
41	ICP-OES	YES	ISO 11885-E22	1,500		1,500			0,90			
29	ISO 12677	YES		1,436	1,405	1,421			0,19			
30	XRF bead			1,490	1,500	1,495			0,85			
31	ISO 12677	YES		1,312	1,291	1,302			0,87			
35	ISO 12677			1,814	1,758	x1,786			3,44	Outlier	x	
37	ICP-OES		Borax fusion	1,260		1,260			1,24			
38	ISO 12677	YES		1,406	1,405	1,406			0,05			
39	XRF bead			1,510	1,630	1,570			1,52			
40	ISO 12677			1,412	1,415	1,414			0,13			
				n	14							
				Mean	1,399							
				Max	1,620							
				Min	1,260							
				Stdev s	0,112							
				C(95%)	0,065			C(95%)=t*s/SQR(n) t(14)=2,160				



Total S expressed as SO3		FLX-CRM 111			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:29:12		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			z>3	n=14 VG=2,371	confirmed	
21	HF-IR			0,153	0,153	0,153			0,38			
22	ISO 12677		reconstitution	0,173	0,180	0,177			0,85			
23	ISO 12677	YES		0,120	0,140	0,130			0,08			
24	HF-IR			0,145	0,140	0,143			0,17			
25	HF-IR			0,150	0,150	0,150			0,32			
26	XRF pellet			0,474	0,826	x0,650			10,29	Outlier	x	
27	HF-IR	YES		0,040	0,033	0,037			1,95			
28	ISO 12677	YES		0,120	0,130	0,125			0,18			
41												
29	ISO 12677	YES		0,064	0,092	0,078			1,12			
30	XRF bead			0,189	0,203	0,196			1,23			
31	ISO 12677	YES		0,161	0,167	0,164			0,60			
35	ISO 12677			0,101	0,098	0,099			0,69			
37	ICP-OES		aqua regia	0,175		0,175			0,82			
38	HF-IR		DIN 51095-1	0,195	0,196	0,196			1,22			
39												
40	ISO 12677			0,073	0,040	0,057			1,55			
				n	14							
				Mean	0,134							
				Max	0,196							
				Min	0,037							
				Stdev s	0,050							
				C(95%)	0,029			C(95%)=t*s/SQR(n) t(14)=2,160				




TiO2		FLX-CRM 111			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:29:21		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean				z>3	n=14 VG=2,371	confirmed
21	ISO 12677			0,158	0,158	0,158				0,29		
22	ISO 12677		reconstitution	0,103	0,109	x0,106				8,99	Outlier	x
23	ISO 12677	YES		0,167	0,173	0,170				1,72		
24	ISO 12677			0,170	0,165	0,168				1,30		
25	XRF bead			0,170	0,160	0,165				0,89		
26	XRF pellet			0,161	0,165	0,163				0,53		
27	ISO 12677	YES		0,160	0,160	0,160				0,05		
28	ISO 12677	YES		0,160	0,160	0,160				0,05		
41	ICP-OES	YES	ISO 11885-E22	0,150		0,150				1,62		
29	ISO 12677	YES		0,165	0,162	0,164				0,63		
30	XRF bead			0,158	0,156	0,157				0,45		
31	ISO 12677	YES		0,188	0,190	x0,189				4,90	Outlier	x
35	ISO 12677			0,149	0,164	0,157				0,54		
37	ICP-OES		Borax fusion	0,250		x0,250				15,11	Outlier	x
38	ISO 12677	YES		0,154	0,154	0,154				0,96		
39	XRF bead			0,150	0,150	0,150				1,62		
40	ISO 12677			0,164	0,159	0,162				0,30		
				n	14							
				Mean	0,160							
				Max	0,170							
				Min	0,150							
				Stdev s	0,006							
				C(95%)	0,003							

C(95%)=t*s/SQR(n) t(14)=2,160

ZrO2		FLX-CRM 111			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:29:26		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean				z>3	n=15 VG=2,409	confirmed
21	XRF pellet			0,047	0,047	0,047				0,85		
22	ISO 12677		reconstitution	0,054	0,044	0,049				0,68		
23	ISO 12677	YES		0,073	0,069	0,071				1,22		
24	ISO 12677			0,051	0,052	0,051				0,47		
25												
26	XRF pellet			0,058	0,055	0,057				0,02		
27	ISO 12677	YES		0,070	0,080	0,075				1,57		
28	ISO 12677	YES		0,060	0,070	0,065				0,71		
41	ICP-OES	YES	ISO 11885-E22	0,060		0,060				0,27		
29	ISO 12677	YES		0,043	0,050	0,047				0,89		
30	XRF bead			0,041	0,044	0,043				1,24		
31	ISO 12677	YES		0,078	0,079	0,078				1,87		
35	ISO 12677			0,042	0,040	0,041				1,37		
37	ICP-OES		Borax fusion	0,054		0,054				0,24		
38	ISO 12677	YES		0,051	0,055	0,053				0,33		
39												
40	ISO 12677			0,067	0,057	0,062				0,45		
				n	15							
				Mean	0,057							
				Max	0,078							
				Min	0,041							
				Stdev s	0,012							
				C(95%)	0,006							


C(95%)=t*s/SQR(n) t(15)=2,145

Cr203		FLX-CRM 111			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:29:57		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			z>3	n=13 VG=2,331	confirmed	
21	XRF pellet			11,940	11,940	11,940						
22	ISO 12677		reconstitution	10,658	10,838	10,748						
23	ISO 12677	YES		11,420	11,500	11,460						
24	ISO 6331			10,777	10,747	10,762						
25	XRF bead			11,690	11,610	11,650						
26	XRF pellet			9,305	9,482	x9,394						
27	ISO 12677	YES		11,560	11,500	11,530						
28	ISO 12677	YES		10,600	10,800	10,700						
41	ICP-OES	YES	ISO 11885-E22	10,000		x10,000						
29	ISO 12677	YES		11,382	10,832	11,107						
30	XRF bead			8,975	9,224	x9,100						
31	ISO 12677	YES		11,950	11,930	11,940						
35	ISO 12677			12,570	12,666	12,618						
37	ICP-OES		Borax fusion	11,250		11,250						
38	ISO 12677	YES		11,701	11,894	11,798						
39												
40	ISO 12677			11,800	11,794	11,797						
				n	13							
				Mean	11,485							
				Max	12,618							
				Min	10,700							
				Stdev s	0,563							
				C(95%)	0,340		C(95%)=t*s/SQR(n) t(13)=2,179					




z-score	Grubbs	Outlier
z>3	n=13 VG=2,331	confirmed
0,81		
1,31		
0,04		
1,28		
0,29		
3,71	Outlier	x
0,08		
1,39		
2,64	Outlier	x
0,67		
4,24	Outlier	x
0,81		
2,01		
0,42		
0,56		
0,55		

Mn304		FLX-CRM 111			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:30:02		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			z>3	n=13 VG=2,331	confirmed	
21	XRF pellet			0,430	0,430	0,430						
22	ISO 12677		reconstitution	0,361	0,364	0,363						
23	ISO 12677	YES		0,280	0,270	0,275						
24	ISO 12677			0,384	0,386	0,385						
25												
26	XRF pellet			0,429	0,439	0,434						
27	ISO 12677	YES		0,340	0,350	0,345						
28	ISO 12677	YES		0,340	0,340	0,340						
41	ICP-OES	YES	ISO 11885-E22	0,320		0,320						
29	ISO 12677	YES		0,408	0,417	0,413						
30	XRF bead			0,686	0,688	x0,687						
31	ISO 12677	YES										
35	ISO 12677			0,444	0,438	0,441						
37	ICP-OES		Borax fusion	0,344		0,344						
38	ISO 12677	YES		0,359	0,358	0,359						
39	XRF bead			0,129	0,129	x0,129						
40	ISO 12677			0,363	0,362	0,363						
				n	13							
				Mean	0,370							
				Max	0,441							
				Min	0,275							
				Stdev s	0,049							
				C(95%)	0,030		C(95%)=t*s/SQR(n) t(13)=2,179					




z-score	Grubbs	Outlier
z>3	n=13 VG=2,331	confirmed
1,23		
0,15		
1,94		
0,30		
1,31		
0,51		
0,61		
1,02		
0,87		
6,48	Outlier	x
1,45		
0,53		
0,23		
4,92	Outlier	x
0,15		

P205		FLX-CRM 111			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:30:11		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			z-score	n=12 VG=2,285	confirmed	
21	ISO 12677			0,082	0,079	0,081			1,02			
22	ISO 12677		reconstitution	0,092	0,087	0,090			0,16			
23	ISO 12677	YES		0,087	0,086	0,087			0,24			
24	ISO 12677			0,099	0,100	0,099			1,45			
25	XRF bead			<0,1	<0,1							
26	XRF pellet			0,222	0,211	x0,217			16,81	Outlier	x	
27	ISO 12677	YES		0,090	0,090	0,090			0,22			
28	ISO 12677	YES		0,080	0,100	0,090			0,22			
41	ICP-OES	YES	ISO 11885-E22	0,090		0,090			0,22			
29	ISO 12677	YES		0,085	0,086	0,086			0,37			
30	XRF bead			0,052	0,056	x0,054			4,48	Outlier	x	
31	ISO 12677	YES		0,346	0,327	x0,337			32,49	Outlier	x	
35	ISO 12677			0,100	0,099	0,099			1,44			
37	ICP-OES		aqua regia	0,071		0,071			2,26	Included		
38	ISO 12677	YES		0,092	0,092	0,092			0,48			
39	XRF bead			0,110	0,120	x0,115			3,49	Outlier	x	
40	ISO 12677			0,087	0,085	0,086			0,30			
				n	12							
				Mean	0,088							
				Max	0,099							
				Min	0,071							
				Stdev s	0,008							
				C(95%)	0,005		C(95%)=t*s/SQR(n) t(12)=2,201					



LOI (@ 1025°C)		FLX-CRM 111			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:30:16		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			z-score	n=13 VG=2,331	confirmed	
21	gravimetric			0,510	0,510	0,510			0,82			
22	ISO 12677			0,635	0,684	0,660			1,04			
23	ISO 12677	YES		0,560	0,540	0,550			0,32			
24	ISO 12677			0,820	0,810	x0,815			2,98	Outlier	x	
25	gravimetric			0,580	0,580	0,580			0,05			
26				0,330	0,310	x0,320			3,18	Outlier	x	
27	ISO 12677	YES		0,700	0,700	0,700			1,55			
28	ISO 12677	YES		0,550	0,560	0,555			0,26			
41												
29	DIN 51081	YES	1 h at 1025°C	0,590	0,610	0,600			0,30			
30	gravimetric			0,592	0,625	0,609			0,41			
31	ISO 12677			0,600	0,600	0,600			0,30			
35	ISO 12677			0,590	0,590	0,590			0,18			
37	gravimetric			0,660		0,660			1,05			
38	ISO 12677	YES		0,440	0,460	0,450			1,56			
39	gravimetric			0,450	0,390	0,420			1,94			
40	ISO 12677			1,500	1,500	x1,500			11,50	Outlier	x	
				n	13							
				Mean	0,576							
				Max	0,700							
				Min	0,420							
				Stdev s	0,080							
				C(95%)	0,049		C(95%)=t*s/SQR(n) t(13)=2,179					



Al2O3		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:47:41				
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean				z-score	Grubbs	Outlier
21	ISO 12677			82,462	82,025	x82,244				>3	n=11 VG=2,234	confirmed
22	ISO 12677		reconstitution	77,563	77,673	x77,618				4,54	Outlier	x
23	ISO 12677	YES		80,020	80,050	80,035				4,10	Outlier	x
24	ISO 12677			79,495	79,594	79,545				0,41		
25	XRF bead	YES		80,340	80,060	80,200				0,50		
26	XRF pellet			55,350	54,190	x54,770				0,72		
27	ISO 12677	YES		80,500	80,900	80,700				46,75	Outlier	x
28	ISO 12677	YES		79,600	79,700	79,650				1,66		
41										0,30		
29	ISO 12677	YES		79,652	79,681	79,667				0,27		
30	XRF bead			79,610	80,860	80,235				0,79		
31	ISO 12677	YES		78,620	78,620	78,620				2,23	Included	
35	ISO 12677			79,832	80,193	80,013				0,37		
37	ICP-OES		Borax fusion	70,990		x70,990				16,47	Outlier	x
38	ISO 12677	YES		79,790	79,842	79,816				0,01		
39	XRF bead			75,230	75,510	x75,370				8,29	Outlier	x
40	ISO 12677			79,574	79,351	79,463				0,65		
				n	11							
				Mean	79,813							
				Max	80,700							
				Min	78,620							
				Stdev s	0,536							
				C(95%)	0,360							

C(95%)=t*s/SQR(n) t(11)=2,228

CaO		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:47:47				
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean				z-score	Grubbs	Outlier
21	ISO 12677			0,205	0,246	x0,226				>3	n=14 VG=2,371	confirmed
22	ISO 12677		reconstitution	0,141	0,154	0,148				2,93	Outlier	x
23	ISO 12677	YES		0,167	0,153	0,160				0,03		
24	ISO 12677			0,165	0,161	0,163				0,49		
25	XRF bead	YES		<0,1	<0,1					0,60		
26	XRF pellet			0,182	0,181	0,182				1,29		
27	ISO 12677	YES		0,110	0,120	0,115				1,17		
28	ISO 12677	YES		0,100	0,100	0,100				1,72		
41	ICP-OES	YES	ISO 11885-E22	0,140		0,140				0,24		
29	ISO 12677	YES		0,124	0,131	0,128				0,71		
30	XRF bead			0,270	0,260	x0,265				4,36	Outlier	x
31	ISO 12677	YES		0,158	0,157	0,157				0,39		
35	ISO 12677			0,150	0,151	0,150				0,14		
37	ICP-OES		aqua regia	0,098		0,098				1,79		
38	ISO 12677	YES		0,178	0,168	0,173				0,97		
39	XRF bead			0,200	0,150	0,175				1,05		
40	ISO 12677			0,169	0,160	0,165				0,66		
				n	14							
				Mean	0,147							
				Max	0,182							
				Min	0,098							
				Stdev s	0,027							
				C(95%)	0,016							

C(95%)=t*s/SQR(n) t(14)=2,160

Fe2O3		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:47:53		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			>3	n=16 VG=2,443	confirmed	
21	ISO 12677			0,338	0,307	0,323						
22	ISO 12677		reconstitution	0,360	0,360	0,360						
23	ISO 12677	YES		0,400	0,380	0,390						
24	ISO 12677			0,379	0,390	0,385						
25	XRF bead	YES		0,400	0,410	0,405						
26	XRF pellet			1,027	1,138	x1,083						
27	ISO 12677	YES		0,390	0,390	0,390						
28	ISO 12677	YES		0,210	0,200	0,205						
41	ICP-OES	YES	ISO 11885-E22	0,340		0,340						
29	ISO 12677	YES		0,399	0,342	0,371						
30	XRF bead			0,240	0,240	0,240						
31	ISO 12677	YES		0,281	0,269	0,275						
35	ISO 12677			0,346	0,350	0,348						
37	ICP-OES		Borax fusion	0,415		0,415						
38	ISO 12677	YES		0,356	0,327	0,342						
39	XRF bead			0,180	0,190	0,185						
40	ISO 12677			0,253	0,226	0,240						
				n	16							
				Mean	0,326							
				Max	0,415							
				Min	0,185							
				Stdev s	0,074							
				C(95%)	0,039							

C(95%)=t*s/SQR(n) t(16)=2,131

HfO2		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:17:35		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			>3	n=12 VG=2,285	confirmed	
21	XRF pellet			0,362	0,362	x0,362						
22	ISO 12677			0,243	0,241	x0,242						
23	ISO 12677	YES		0,121	0,122	0,122						
24	ISO 12677			0,138	0,139	0,139						
25	XRF bead			0,130	0,120	0,125						
26	XRF pellet			0,082	0,087	0,085						
27	ICP-OES	YES		0,117	0,120	0,119						
28	ISO 12677	YES		0,070	0,070	0,070						
41	ICP-OES	YES	ISO 11885-E22	0,044		0,044						
29	ISO 12677	YES		0,100	0,097	0,099						
30												
31	ISO 12677	YES		0,107	0,107	0,107						
35	ISO 12677			0,082	0,081	0,082						
37												
38	ISO 12677	YES		0,070	0,079	0,075						
39												
40	ISO 12677			0,124	0,124	0,124						
				n	12							
				Mean	0,099							
				Max	0,139							
				Min	0,044							
				Stdev s	0,028							
				C(95%)	0,018							

C(95%)=t*s/SQR(n) t(12)=2,201

K2O		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:48:04		
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean	z-score	Grubbs	Outlier	
21	ISO 12677			0,092	0,092	0,092	>3	n=10 VG=2,176	confirmed	
22	ISO 12677		reconstitution	0,075	0,106	0,091	0,35			
23	ISO 12677	YES		0,087	0,081	0,084	0,07			
24	ISO 12677			0,099	0,094	0,097	1,16			
25	XRF bead	YES		<0,1	<0,1		1,20			
26	XRF pellet			0,176	0,172	x0,174	15,80	Outlier	x	
27	ICP-OES	YES		0,095	0,095	0,095	0,91			
28	ISO 12677	YES		0,080	0,090	0,085	0,97			
41	ICP-OES	YES	ISO 11885-E22	0,040		x0,040	9,45	Outlier	x	
29	ISO 12677	YES		0,100	0,098	0,099	1,67			
30	XRF bead			0,177	0,190	x0,184	17,68	Outlier	x	
31	ISO 12677	YES		0,051	0,052	x0,051	7,38	Outlier	x	
35	ISO 12677			0,071	0,074	x0,073	3,23	Outlier	x	
37	ICP-OES		aqua regia	0,048		x0,048	7,94	Outlier	x	
38	ISO 12677	YES		0,089	0,085	0,087	0,59			
39	XRF bead			0,090	0,080	0,085	0,97			
40	ISO 12677			0,089	0,086	0,088	0,50			
				n	10					
				Mean	0,090					
				Max	0,099					
				Min	0,084					
				Stdev s	0,005					
				C(95%)	0,004					

C(95%)=t*s/SQR(n) t(10)=2,262

MgO		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:48:08		
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean	z-score	Grubbs	Outlier	
21	ISO 12677			0,789	0,808	0,799	>3	n=16 VG=2,443	confirmed	
22	ISO 12677		reconstitution	0,901	0,883	0,892	0,55			
23	ISO 12677	YES		0,720	0,720	0,720	1,72			
24	ISO 12677			0,651	0,630	0,641	0,44			
25	XRF bead	YES		0,690	0,680	0,685	1,43			
26	XRF pellet			0,864	0,838	0,851	0,87			
27	ISO 12677	YES		0,570	0,540	x0,555	1,21			
28	ISO 12677	YES		0,660	0,630	0,645	2,50	Outlier	x	
41	ICP-OES	YES	ISO 11885-E22	0,690		0,690	1,38			
29	ISO 12677	YES		0,781	0,815	0,798	0,81			
30	XRF bead			0,824	0,844	0,834	0,54			
31	ISO 12677	YES		0,819	0,833	0,826	0,99			
35	ISO 12677			0,779	0,715	0,747	0,89			
37	ICP-OES		aqua regia	0,630		0,630	0,10			
38	ISO 12677	YES		0,781	0,782	0,782	1,56			
39	XRF bead			0,760	0,770	0,765	0,34			
40	ISO 12677			0,772	0,771	0,772	0,13			
				n	16					
				Mean	0,755					
				Max	0,892					
				Min	0,630					
				Stdev s	0,080					
				C(95%)	0,042					

C(95%)=t*s/SQR(n) t(16)=2,131

Na2O		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:48:13		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean				>3	n=14 VG=2,371	confirmed
21	ISO 12677			0,348	0,368	0,358				1,38		
22	ISO 12677		reconstitution	0,184	0,191	0,188				1,20		
23	ISO 12677	YES		0,330	0,330	0,330				0,96		
24												
25	ICP-OES	YES		0,260	0,240	0,250				0,25		
26	XRF pellet			1,233	1,158	x1,196				14,04	Outlier	x
27	ICP-OES	YES		0,310	0,310	0,310				0,66		
28	ISO 12677	YES		0,180	0,170	0,175				1,38		
41	ICP-OES	YES	ISO 11885-E22	0,230		0,230				0,55		
29	ISO 12677	YES		0,312	0,301	0,307				0,60		
30	XRF bead			0,277	0,266	0,272				0,07		
31	ISO 12677	YES		0,628	0,586	x0,607				5,14	Outlier	x
35	ISO 12677			0,259	0,272	0,266				0,02		
37	ICP-OES		aqua regia	0,135		0,135				1,99		
38	ISO 12677	YES		0,316	0,317	0,317				0,75		
39	XRF bead			0,260	0,260	0,260				0,10		
40	ISO 12677			0,332	0,342	0,337				1,06		
				n	14							
				Mean	0,267							
				Max	0,358							
				Min	0,135							
				Stdev s	0,066							
				C(95%)	0,038							

C(95%)=t*s/SQR(n) t(14)=2,160

NiO		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:31:40		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean				>3	n=10 VG=2,176	confirmed
21	XRF pellet			0,019	0,019	0,019				1,34		
22	ISO 12677		reconstitution	0,014	0,015	0,015				0,71		
23	ISO 12677	YES		0,001	0,001	0,001				1,16		
24	ISO 12677			<0,005	<0,005							
25												
26	XRF pellet			0,002	0,003	0,002				0,96		
27	ICP-OES	YES		0,002	0,002	0,002				1,04		
28	ISO 12677	YES		0,010	0,010	0,010				0,09		
41	ICP-OES	YES	ISO 11885-E22	<0,005								
29	ISO 12677	YES		0,013	0,015	0,014				0,64		
30												
31	ISO 12677	YES		0,007	0,007	0,007				0,28		
35	ISO 12677			<0,004	<0,004							
37												
38	ISO 12677	YES		0,020	0,020	0,020				1,48		
39												
40	ISO 12677			0,004	0,003	0,004				0,81		
				n	10							
				Mean	0,009							
				Max	0,020							
				Min	0,001							
				Stdev s	0,007							
				C(95%)	0,005							

C(95%)=t*s/SQR(n) t(10)=2,262

SiO2		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:48:20		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			>3	n=14 VG=2,371	confirmed	
21	ISO 12677			12,410	12,810	12,610	21		1,75			
22	ISO 12677		reconstitution	12,357	12,427	12,392	22		0,90			
23	ISO 12677	YES		11,920	11,880	11,900	23		1,04			
24	ISO 12677			11,990	12,050	12,020	24		0,57			
25	XRF bead	YES		11,880	11,910	11,895	25		1,06			
26	XRF pellet			17,280	17,030	x17,155	26		19,61	Outlier	x	
27	ISO 12677	YES		12,270	12,060	12,165	27		0,00			
28	ISO 12677	YES		12,200	12,100	12,150	28		0,06			
41							41					
29	ISO 12677	YES		12,403	12,392	12,398	29		0,92			
30	XRF bead			11,810	11,710	11,760	30		1,59			
31	ISO 12677	YES		12,420	12,370	12,395	31		0,91			
35	ISO 12677			12,223	12,104	12,164	35		0,00			
37	ICP-OES		Borax fusion	12,020		12,020	37		0,57			
38	ISO 12677	YES		12,043	11,875	11,959	38		0,81			
39	XRF bead			11,530	11,420	x11,475	39		2,71	Outlier	x	
40	ISO 12677			12,511	12,430	12,471	40		1,20			
				n	14							
				Mean	12,164							
				Max	12,610							
				Min	11,760							
				Stdev s	0,254							
				C(95%)	0,147							

C(95%)=t*s/SQR(n) t(14)=2,160


Total S expressed as SO3		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:48:27		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			>3	n=10 VG=2,176	confirmed	
21	HF-IR			0,077	0,077	0,077	21		1,21			
22	ISO 12677		reconstitution	0,086	0,088	0,087	22		1,57			
23	ISO 12677	YES		<0,03	<0,03		23					
24							24					
25	HF-IR			0,010	0,030	0,020	25		0,85			
26	XRF pellet			0,041	0,059	0,050	26		0,25			
27	HF-IR	YES		0,010	0,015	0,013	27		1,12			
28	ISO 12677	YES		<0,01	<0,01		28					
41							41					
29	ISO 12677	YES		0,042	0,064	0,053	29		0,34			
30	XRF bead			0,065	0,065	0,065	30		0,78			
31	ISO 12677	YES		0,040	0,045	0,042	31		0,05			
35	ISO 12677			<0,010	<0,010		35					
37							37					
38	HF-IR		DIN 51095-1	0,016	0,011	0,014	38		1,09			
39							39					
40	ISO 12677			0,012	0,017	0,015	40		1,05			
				n	10							
				Mean	0,044							
				Max	0,087							
				Min	0,013							
				Stdev s	0,028							
				C(95%)	0,020							

C(95%)=t*s/SQR(n) t(10)=2,262

TiO2		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:48:32		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			>3	n=16 VG=2,443	confirmed	
21	ISO 12677			0,277	0,276	0,277	21		0,15			
22	ISO 12677		reconstitution	0,257	0,257	0,257	22		0,64			
23	ISO 12677	YES		0,200	0,200	x0,200	23		2,94	Outlier	x	
24	ISO 12677			0,297	0,304	0,301	24		1,12			
25	XRF bead	YES		0,280	0,290	0,285	25		0,49			
26	XRF pellet			0,315	0,307	0,311	26		1,53			
27	ISO 12677	YES		0,290	0,300	0,295	27		0,90			
28	ISO 12677	YES		0,270	0,270	0,270	28		0,11			
41	ICP-OES	YES	ISO 11885-E22	0,250		0,250	41		0,92			
29	ISO 12677	YES		0,280	0,271	0,276	29		0,11			
30	XRF bead			0,223	0,222	0,223	30		2,03			
31	ISO 12677	YES		0,248	0,252	0,250	31		0,92			
35	ISO 12677			0,273	0,267	0,270	35		0,11			
37	ICP-OES		Borax fusion	0,317		0,317	37		1,78			
38	ISO 12677	YES		0,277	0,267	0,272	38		0,03			
39	XRF bead			0,260	0,250	0,255	39		0,72			
40	ISO 12677			0,261	0,255	0,258	40		0,60			
				n	16							
				Mean	0,273							
				Max	0,317							
				Min	0,223							
				Stdev s	0,025							
				C(95%)	0,013							
						C(95%)=t*s/SQR(n)	t(16)=2,131					


ZrO2		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:48:42		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			>3	n=13 VG=2,331	confirmed	
21	XRF pellet			2,208	2,208	x2,208	21		11,23	Outlier	x	
22	ISO 12677		reconstitution	5,706	5,951	5,829	22		0,36			
23	ISO 12677	YES		5,830	5,880	5,855	23		0,28			
24	ISO 12677			5,704	5,715	5,710	24		0,72			
25	XRF bead			6,010	5,990	6,000	25		0,15			
26	XRF pellet			5,745	5,720	5,733	26		0,65			
27	ISO 12677	YES		6,300	6,100	6,200	27		0,75			
28	ISO 12677	YES		6,470	6,460	6,465	28		1,55			
41							41					
29	ISO 12677	YES		5,573	5,593	5,583	29		1,10			
30	XRF bead			0,950	0,942	x0,946	30		15,02	Outlier	x	
31	ISO 12677	YES		6,518	6,605	6,562	31		1,84			
35	ISO 12677			6,005	5,927	5,966	35		0,05			
37	ICP-OES		Borax fusion	5,360		5,360	37		1,77			
38	ISO 12677	YES		5,962	6,126	6,044	38		0,28			
39							39					
40	ISO 12677			6,045	6,022	6,034	40		0,25			
				n	13							
				Mean	5,949							
				Max	6,562							
				Min	5,360							
				Stdev s	0,333							
				C(95%)	0,201							
						C(95%)=t*s/SQR(n)	t(13)=2,179					

Co304		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:46:51				
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean				z-score	Grubbs	Outlier
21	XRF pellet			0,013	0,013	0,013				>3	n=7 VG=1,938	confirmed
22	ISO 12677			0,007	0,007	0,007				1,07		
23	ISO 12677	YES		0,006	0,007	0,007				0,35		
24	ISO 12677			<0,02	<0,02					0,55		
25												
26	XRF pellet			0,014	0,015	0,014				1,44		
27	ICP-OES	YES		0,005	0,005	0,005				0,89		
28	ISO 12677	YES		<0,01	<0,01							
41	ICP-OES	YES	ISO 11885-E22	0,024		x0,024				3,81	Outlier	x
29												
30												
31	ISO 12677	YES										
35	ISO 12677			<0,005	<0,005							
37	ICP-OES		aqua regia	0,004		0,004				1,17		
38	ISO 12677	YES		<0,01	<0,01							
39												
40	ISO 12677			0,011	0,010	0,011				0,45		
				n	7							
				Mean	0,009							
				Max	0,014							
				Min	0,004							
				Stdev s	0,004							
				C(95%)	0,004							



C(95%)=t*s/SQR(n) t(7)=2,447

WO3		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:47:17				
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean				z-score	Grubbs	Outlier
21	XRF pellet			0,070	0,070	0,070				>3	n=10 VG=2,176	confirmed
22	ISO 12677		reconstitution	0,043	0,041	0,042				1,41		
23	ISO 12677	YES		0,013	0,013	0,013				0,07		
24										1,32		
25												
26	XRF pellet			0,035	0,036	0,035				0,24		
27	ICP-OES	YES		0,034	0,037	0,036				0,24		
28	ISO 12677	YES		0,040	0,040	0,040				0,03		
41												
29												
30												
31	ISO 12677	YES		0,043	0,043	0,043				0,11		
35	ISO 12677			0,033	0,031	0,032				0,41		
37	ICP-OES		aqua regia	0,517		x0,517				22,80	Outlier	x
38	ISO 12677	YES		0,010	0,020	0,015				1,22		
39												
40	ISO 12677			0,076	0,083	0,080				1,86		
				n	10							
				Mean	0,041							
				Max	0,080							
				Min	0,013							
				Stdev s	0,021							
				C(95%)	0,015							



C(95%)=t*s/SQR(n) t(10)=2,262

Cr2O3		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:48:51		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			z>3	n=10 VG=2,176	confirmed	
21	XRF pellet			0,130	0,130	x0,130						
22	ISO 12677		reconstitution	0,011	0,011	0,011						
23	ISO 12677	YES		0,020	0,021	0,021						
24	ISO 12677			0,021	0,022	0,022						
25												
26	XRF pellet			0,242	0,295	x0,269						
27	ISO 12677	YES		0,030	0,020	0,025						
28	ISO 12677	YES		<0,01	<0,01							
41	ICP-OES	YES	ISO 11885-E22	0,014		0,014						
29	ISO 12677	YES		0,022	0,024	0,023						
30	XRF bead			0,013	0,012	0,013						
31	ISO 12677	YES		0,016	0,018	0,017						
35	ISO 12677			0,013	0,017	0,015						
37	ICP-OES		Borax fusion	0,015		0,015						
38	ISO 12677	YES		0,059	0,057	x0,058						
39												
40	ISO 12677											
						n	10					
						Mean	0,017					
						Max	0,025					
						Min	0,011					
						Stdev s	0,005					
						C(95%)	0,003					

C(95%)=t*s/SQR(n) t(10)=2,262

Mn3O4		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:48:57		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean			z>3	n=12 VG=2,285	confirmed	
21	XRF pellet			0,034	0,034	0,034						
22	ISO 12677		reconstitution	0,026	0,024	0,025						
23	ISO 12677	YES		0,028	0,025	0,027						
24	ISO 12677			<0,03	<0,03							
25												
26	XRF pellet			0,026	0,028	0,027						
27	ISO 12677	YES		0,030	0,030	0,030						
28	ISO 12677	YES		0,020	0,020	0,020						
41	ICP-OES	YES	ISO 11885-E22	0,014		0,014						
29	ISO 12677	YES		0,029	0,033	0,031						
30	XRF bead			0,043	0,045	0,044						
31	ISO 12677	YES										
35	ISO 12677			<0,010	<0,010							
37	ICP-OES		Borax fusion	0,014		0,014						
38	ISO 12677	YES		0,020	0,010	0,015						
39	XRF bead			0,010	0,000	0,010						
40	ISO 12677											
						n	12					
						Mean	0,024					
						Max	0,044					
						Min	0,010					
						Stdev s	0,010					
						C(95%)	0,006					

C(95%)=t*s/SQR(n) t(12)=2,201

P205		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:49:02		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean						
21	ISO 12677			0,069	0,068	0,069				>3	n=10 VG=2,176	confirmed
22	ISO 12677		reconstitution	0,126	0,123	x0,125				0,43		
23	ISO 12677	YES		0,082	0,082	0,082				4,34	Outlier	x
24	ISO 12677			0,081	0,081	0,081				0,71		
25	XRF bead	YES		<0,1	<0,1					0,62		
26	XRF pellet			0,230	0,222	x0,226				12,87	Outlier	x
27	ISO 12677	YES		0,080	0,080	0,080				0,54		
28	ISO 12677	YES		0,070	0,070	0,070				0,30		
41	ICP-OES	YES	ISO 11885-E22	0,061		0,061				1,06		
29	ISO 12677	YES		0,130	0,136	x0,133				5,01	Outlier	x
30	XRF bead			0,037	0,039	x0,038				3,01	Outlier	x
31	ISO 12677	YES		<0.00069	<0.00069							
35	ISO 12677			0,088	0,088	0,088				1,22		
37	ICP-OES		aqua regia	0,053		0,053				1,74		
38	ISO 12677	YES		0,089	0,086	0,088				1,17		
39	XRF bead			0,070	0,060	0,065				0,73		
40	ISO 12677											
						n	10					
						Mean	0,074					
						Max	0,088					
						Min	0,053					
						Stdev s	0,012					
						C(95%)	0,008					

C(95%)=t*s/SQR(n) t(10)=2,262

LOI (@ 1025°C)		FLX-CRM 112			Mass %	Mass %	Mass %	Freitag, 12. Oktober 2012 11:49:06		z-score	Grubbs	Outlier
Lab.No:	Method	ISO 17025	Remark	Meas #1	Meas #2	Mean						
21	gravimetric			5,350	5,200	5,275				>3	n=13 VG=2,331	confirmed
22	ISO 12677			5,559	5,559	5,559				1,18		
23	ISO 12677	YES		5,020	4,990	x5,005				1,07		
24	ISO 12677			5,200	5,270	5,235				3,32	Outlier	x
25	gravimetric			4,220	4,240	x4,230				1,50		
26	ISO 12677			5,260	5,280	5,270				9,46	Outlier	x
27	ISO 12677	YES		5,660	5,660	5,660				1,22		
28	ISO 12677	YES		5,400	5,430	5,415				1,87		
41										0,07		
29	DIN 51081	YES		5,370	5,360	5,365				0,47		
30	gravimetric			5,441	5,533	5,487				0,50		
31	ISO 12677			5,400	5,400	5,400				0,19		
35	ISO 12677			5,380	5,380	5,380				0,35		
37	gravimetric			5,480		5,480				0,45		
38												
39	950°C			5,390	5,430	5,410				0,11		
40	ISO 12677			5,657	5,488	5,573				1,18		
						n	13					
						Mean	5,424					
						Max	5,660					
						Min	5,235					
						Stdev s	0,126					
						C(95%)	0,076					

C(95%)=t*s/SQR(n) t(13)=2,179